

# LCD - Corneal Hysteresis (L38211)

Links in PDF documents are not guaranteed to work. To follow a web link, please use the MCD Website.

## Contractor Information

CONTRACTOR NAME	CONTRACT TYPE	CONTRACT NUMBER	JURISDICTION	STATES
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part A	05101 - MAC A	J - 05	Iowa
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part B	05102 - MAC B	J - 05	Iowa
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part A	05201 - MAC A	J - 05	Kansas
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part B	05202 - MAC B	J - 05	Kansas
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part A	05301 - MAC A	J - 05	Missouri - Entire State
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part B	05302 - MAC B	J - 05	Missouri - Entire State
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part A	05401 - MAC A	J - 05	Nebraska
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part B	05402 - MAC B	J - 05	Nebraska
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part A	05901 - MAC A	J - 05	Alabama Alaska Arizona Arkansas California - Entire State Colorado Connecticut Delaware Florida Georgia Hawaii Idaho Illinois Indiana Iowa Kansas Kentucky Louisiana

CONTRACTOR NAME	CONTRACT TYPE	CONTRACT NUMBER	JURISDICTION	STATES
				Maine Maryland Massachusetts Michigan Mississippi Missouri - Entire State Montana Nebraska Nevada New Hampshire New Jersey New Mexico North Carolina North Dakota Ohio Oklahoma Oregon Pennsylvania Rhode Island South Carolina South Dakota Tennessee Texas Utah Vermont Virginia Washington West Virginia Wisconsin Wyoming
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part A	08101 - MAC A	J - 08	Indiana
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part B	08102 - MAC B	J - 08	Indiana
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part A	08201 - MAC A	J - 08	Michigan
<a href="#">Wisconsin Physicians Service Insurance Corporation</a>	MAC - Part B	08202 - MAC B	J - 08	Michigan

## LCD Information

# Document Information

**LCD ID**

L38211

**LCD Title**

Corneal Hysteresis

**Proposed LCD in Comment Period**

N/A

**Source Proposed LCD**

[DL38211](#)

**Original Effective Date**

For services performed on or after 10/14/2019

**Revision Effective Date**

For services performed on or after 07/27/2023

**Revision Ending Date**

N/A

**Retirement Date**

N/A

**Notice Period Start Date**

08/29/2019

**Notice Period End Date**

10/13/2019

## Issue

**Issue Description**

Biannual review completed. Minor grammatical changes made throughout. Updated references to include superscripts throughout the LCD. Deleted bibliography references for LCDs and LCAs that have been retired, updated references in bibliography and fixed broken hyperlinks in bibliography. No change in coverage.

## CMS National Coverage Policy

N/A

## Coverage Guidance

**Coverage Indications, Limitations, and/or Medical Necessity**

This is a NON-coverage policy for all corneal hysteresis assessments as a means of risk assessment or monitoring for

## AMA CPT / ADA CDT / AHA NUBC Copyright Statement

CPT codes, descriptions and other data only are copyright 2022 American Medical Association. All Rights Reserved. Applicable FARS/HHSARS apply.

Fee schedules, relative value units, conversion factors and/or related components are not assigned by the AMA, are not part of CPT, and the AMA is not recommending their use. The AMA does not directly or indirectly practice medicine or dispense medical services. The AMA assumes no liability for data contained or not contained herein.

Current Dental Terminology © 2022 American Dental Association. All rights reserved.

Copyright © 2023, the American Hospital Association, Chicago, Illinois. Reproduced with permission. No portion of the American Hospital Association (AHA) copyrighted materials contained within this publication may be copied without the express written consent of the AHA. AHA copyrighted materials including the UB04 codes and descriptions may not be removed, copied, or utilized within any software, product, service, solution or derivative work without the written consent of the AHA. If an entity wishes to utilize any AHA materials, please contact the AHA at 312□893□6816.

Making copies or utilizing the content of the UB04 Manual, including the codes and/or descriptions, for internal purposes, resale and/or to be used in any product or publication; creating any modified or derivative work of the UB04 Manual and/or codes and descriptions; and/or making any commercial use of UB04 Manual or any portion thereof, including the codes and/or descriptions, is only authorized with an express license from the American Hospital Association. The American Hospital Association (the "AHA") has not reviewed, and is not responsible for, the completeness or accuracy of any information contained in this material, nor was the AHA or any of its affiliates, involved in the preparation of this material, or the analysis of information provided in the material. The views and/or positions presented in the material do not necessarily represent the views of the AHA. CMS and its products and services are not endorsed by the AHA or any of its affiliates.

progression of ophthalmic disease activity.

## Summary of Evidence

Hysteresis is a measure of resistance to deformation to an applied force. Corneal hysteresis (CH) is a measure of the viscoelastic dampening property of the cornea and is postulated to be a surrogate for the viscoelastic dampening properties of the posterior sclera and lamina cribrosa through which the retinal ganglion cell axons pass as they exit the eye. It has been theorized that glaucomatous damage to the retinal ganglion cell axons occurs at the lamina cribrosa and that viscoelastic differences in the lamina cribrosa are responsible for differential effects of intraocular pressure within these tissues and contribute to the susceptibility to intraocular pressure (IOP)-mediated damage. Studies show an association between a lower CH and glaucoma or glaucoma risk, and it has been proposed as a risk stratification tool for use in the treatment of glaucoma, glaucoma suspect, and ocular hypertension. Corneal hysteresis is not itself a modifiable risk factor for glaucoma, but theoretically could signal the need for more aggressive IOP reduction.

The Ocular Response Analyzer (ORA) (Reichert Inc., Buffalo, New York, USA) is a non-contact tonometer that measures CH. The ORA received FDA 510(k) clearance in 2004 for the intended use of measurement of IOP and biomechanical response of the cornea "for the purpose of aiding in the diagnosis and monitoring of glaucoma."<sup>1</sup> This device measures CH by measuring the difference of 2 applanation event pressures taken during the inward and outward movement of the cornea following delivery of a metered pulse of air.

A 2006 retrospective, observational study compared CH and central corneal thickness (CCT) on various indices of glaucomatous damage in 230 patients (mean 65 years), 85% diagnosed with primary open angle glaucoma (POAG) or glaucoma suspect, and 15% with angle-closure glaucoma, suspected angle-closure glaucoma, or secondary glaucoma.<sup>2</sup> Multivariate analysis found CCT, but not CH to be predictive of higher cup-to-disk ratio (CDR) ( $P=0.02$  vs  $P=0.36$ ). Multivariate analysis found lower CH but not CCT to be predictive of visual field progression ( $P = 0.30$ ), but not after factoring in axial length ( $P = 0.09$ ). Neither CH nor CCT were significantly associated with worsening mean deviation (MD) or pattern standard deviation (PSD). A 2011 prospective observational study of 162 POAG subjects found no statistical difference in CH compared with 150 normal subjects.<sup>3</sup> A small (57 patients), 2012 retrospective study found both CH and IOP to be independent, statistically significant predictors of response to topical prostaglandin treatment.<sup>4</sup>

A 2012 prospective cohort study of 153 patients (153 eyes) with established glaucoma evaluated the relationship between CCT and CH and their correlation with progressive visual field (VF) loss.<sup>5</sup> Baseline measurements included age, race, sex, CH, MD, PSD, CCT, and IOP (calculated by averaging the first 4 measurements following the baseline VF), peak IOP, and corneal compensated IOP (IOPcc). Progression of glaucoma was determined by an automated pointwise linear regression analysis of visual field tests. Progression occurred in 25 enrolled eyes (16%) and demonstrated significantly lower CCT and CH compared with non-progressed eyes ( $p=0.04$  and  $p<0.01$ , respectively). There was significant correlation between CH and CCT ( $r=0.33$ ,  $P<0.01$ ). After multivariate analysis, peak IOP, age, and CH were demonstrated to be significantly associated with glaucomatous visual field progression. The authors conclude that "as CH may describe corneal properties more completely than thickness alone, it may be a parameter that is better associated with progression." However, of the 25 subjects that demonstrated glaucomatous progression, 9 had either secondary glaucoma, juvenile glaucoma, or angle closure glaucoma. No subgroup analysis was performed. Other confounders include the use of CH measurements obtained during a non-standardized episode in the care continuum (not at baseline), as well as the non-standardized treatment and follow-up (provider discretion) protocol.

A 2013 retrospective study of 131 glaucoma patients investigated the correlation between CH and other structural markers of glaucomatous damage on spectral domain optical coherence tomography (SDOCT).<sup>6</sup> In a multivariable analysis including MD, age, average retinal nerve fiber layer (RNFL) thickness, and glaucoma status, only MD ( $p =$

0.001) and age ( $p < 0.001$ ) retained significant associations with CH. The authors conclude that “in patients under evaluation and treatment for glaucoma, CH was more closely related to visual field MD than to structural markers of glaucoma damage as measured by SDOCT.”

A 2017 cross sectional study compared single CH measurements among 123 patients (123 eyes) previously diagnosed with either glaucoma (high tension glaucoma,  $N=37$ ; pseudoexfoliative glaucoma,  $N=12$ ; normal tension glaucoma,  $N=24$ ), ocular hypertension (OHT) ( $N=28$ ), or glaucoma-like optic discs (GLD) ( $N=22$ ).<sup>7</sup> A One-way Analysis of Covariance (ANCOVA), correcting for differences in age and IOP, found mean CH to be significantly lower in patients with glaucoma versus those with OHT and GLD ( $p < 0.001$ ). The authors hypothesize there may be greater viscoelasticity in ocular tissues of GLD and OHT which may have a protective role against glaucomatous nerve damage.

The following 4 studies by the same principal investigator included subjects who were part of the larger Diagnostic Innovations in Glaucoma Study (DIGS).<sup>8-11</sup> DIGS is a single-center, prospective, longitudinal cohort study of the relationships between optic nerve structure and glaucomatous vision loss, and the assessment of new diagnostic and monitoring modalities that could be used to mitigate functional vision loss by identifying at-risk patients through earlier detection and intervention.

A 2012 observational cross-sectional study of the association between CH and severity of glaucoma, as defined by automated visual field deficits and RNFL thickness, among 299 eyes in 191 glaucoma or glaucoma suspect patients.<sup>8</sup> In multivariable regression models, after adjusting for central corneal thickness, age, and axial length, the relationship of CH to RNFL thickness was not statistically significant. The authors conclude they found only “a weak relationship between corneal biomechanical parameters and measures of structural and functional damage in glaucoma.”

A 2013 prospective, observational study looked at the relationship between baseline CH and visual field progression in 68 patients (114 eyes) with confirmed diagnosis of open angle glaucoma.<sup>9</sup> CH measurements were obtained at the baseline study visit. Subjects underwent baseline and every 6 month follow-up examinations which included examination and assessment of Goldmann applanation tonometer (GAT) IOP, CCT, Humphrey visual field, stereoscopic disc photos, and axial length measurements. Treatment was uncontrolled and at the discretion of the treating physician. Subjects were followed for an average of 4 years (range, 2.0-6.6 years), during which visual fields were assessed for evidence of progression using the visual field index (VFI) method. Univariable analysis found that each 1 mmHg lower baseline CH was significantly associated with a 0.25%/year faster rate of visual field progression ( $p < 0.001$ ). The multivariable model showed an interaction between IOP and CH; eyes with high IOP and low CH were at increased risk for having fast rates of disease progression. CH explained a larger proportion of the variation in VFI change than CCT (17.4% vs. 5.2%, respectively). The authors conclude: “The prospective longitudinal design of this study supports the role of CH as an important factor to be considered in the assessment of the risk of progression in glaucoma patients.”

In a 2016 prospective, observational cohort study, the relationship between CH and progressive loss of the retinal nerve fiber layer (RNFL) was analyzed in 133 patients (186 eyes) with confirmed diagnosis of open angle glaucoma.<sup>10</sup> CH measurements were obtained at the baseline study visit. Subjects underwent baseline and every 6 month follow-up examinations which included examination and assessment of GAT IOP, CCT, Humphrey visual field, stereoscopic disc photos, and circumpapillary retinal nerve fiber layer thickness measurements with the spectral domain optical coherence tomography (SD-OCT). Treatment was uncontrolled and at the discretion of the treating physician. Subjects were followed for an average of 3.8 years (range, 2.0-5.2 years) during which time average circumpapillary RNFL thickness measurements and stereodisc photos were assessed for evidence of glaucomatous progression. Univariable analysis found that each 1 mmHg lower baseline CH was significantly associated with a 0.13  $\mu\text{m}/\text{year}$  faster loss of RNFL ( $p = 0.011$ ). In multivariable analysis adjusting for age, race, average GAT IOP and CCT, CH was still associated with a faster rate of RNFL loss ( $p=0.015$ ). The authors conclude that “the prospective

longitudinal design of this study supports a role for CH as a risk factor for progression in glaucoma.” Both the Medeiros and Zhang studies were small and confounded by the fact that treatment was not controlled. Though findings were suggestive, the use of a complex regression model that was not clearly developed from a-priori hypothesized relationship and not validated following development do not allow firm conclusions about the generalizability of the results.

A 2018 prospective, observational study investigated the predictive role of CH as a risk factor for the development of glaucoma in a cohort of glaucoma suspect patients.<sup>11</sup> The study included 199 patients (287 eyes) recruited from a single site. Treatment for glaucoma suspect was uncontrolled and subject to discretion of the treating physician. Baseline measurements included CH, GAT IOP, CCT, Humphrey visual field, and stereoscopic disc examination. Subjects were examined every 6 months for an average follow-up period of 3.9 years during which time glaucoma developed in 19% of enrolled eyes (54 eyes in 48 patients). Baseline CH and age was significantly lower in those who developed glaucoma vs those who did not ( $9.5 \pm 1.5$  mm Hg vs.  $10.2 \pm 2.0$  mm Hg;  $p = 0.012$ ). Baseline MD and PSD were significantly different between the two groups. CH was found to be predictive of glaucoma development in a multivariable model (hazard ratio = 1.20; 95% CI: 1.01-1.42;  $p = 0.04$ ), while baseline IOP, CCT, and treatment were not. Each 1 mmHg lower CH was associated with a 21% increase in risk of glaucoma development (95% CI 1.04-1.41;  $p = 0.013$ ). The authors acknowledged that “because the impact of CCT on risk of glaucoma development is now widely known, it is likely that physicians may have treated more aggressively eyes of glaucoma suspects who had thin corneas, also artificially reducing the impact of CCT as a predictive factor,” and that, “the higher predictive value of CH compared to CCT in our study should be seen with caution.” Additionally, the multivariable analysis included only some of the known risk factors of glaucoma development, specifically, age, IOP, CCT, PSD, and treatment, but excluded others such as race, family history, and optic disc morphology (CDR). The authors conclude that “future studies including randomization protocols controlling for treatment should be performed to clarify the relative importance of these predictive factors.”

A 2017 meta-analysis included 19 studies that assessed CH in 1213 eyes with glaucoma and 1055 healthy eyes.<sup>12</sup> Mean CH was 1.5 mm Hg lower, and mean CCT 8.5 micrometer less thick, in eyes with glaucoma ( $P < 0.0001$  and  $P < 0.001$ , respectively). The authors conclude that there are differences in corneal properties such as CH and CCT between patients with glaucoma and healthy controls “and support further studies on the influence of CH and CCT in glaucoma screening and diagnosis.” In a 2018 prospective cross-sectional study of CH as a potential glaucoma screening tool in 46 patients (76 eyes) on routine eye exam, the 21 eyes (27.6%) found to have normal tension glaucoma (NTG) did not differ statistically in CH ( $P = 0.19$ ).<sup>13</sup>

A limited body of evidence suggests there may be a role in the application of CH in the identification of corneal pathology or in preoperative assessment prior to refractive surgery. A 2007 study first described a statistically significant difference in the mean CH of 207 normal and 93 keratoconic eyes ( $10.7 + 2.0$  mmHg vs.  $9.6 + 2.2$  mmHg;  $p < 0.0001$ ).<sup>14</sup> The study also revealed that CH values in the keratoconic eyes decreased with increasing severity of disease, though could not differentiate between eyes with mild keratoconus and normal controls. A subsequent study similarly found poor overall predictive accuracy for CH to detect mild keratoconus from age- and gender-matched controls.<sup>15</sup> A 2011 study investigated the ability of the ORA parameters to aid in diagnosis of keratoconus in preoperative laser in situ keratomileusis (LASIK) patients.<sup>16</sup> Biomechanical measurements were acquired from 103 eyes with mild keratoconus and 97 control eyes and 12 parameters were analyzed. Though sensitivity and specificity of the parameters was low (66% and 67%, respectively, for CH), the authors concluded that some parameters offered high negative likelihood ratios and should be studied in a larger sample size.

### **Analysis of Evidence (Rationale for Determination)**

Neither the current (2015) nor most recent updates (2017) to the American Academy of Ophthalmology (AAO) Preferred Practice Pattern (PPP) guidelines for glaucoma recommend measurement of CH in the management or risk assessment of glaucoma, glaucoma suspect, or ocular hypertension.<sup>17-19</sup> Similarly, the Canadian Association of Optometrists (CAO), notes that “despite the association between CH and glaucoma onset and progression, there is

still a paucity of clinical evidence to support adding CH measurement to the standard glaucoma workup.<sup>20</sup> The AAO PPP 2018 guidelines for corneal ectasia concede that while measures of corneal biomechanics, including CH, are likely altered in corneal ectasia, the parameters for use in the detection at the subclinical stage is currently being evaluated.<sup>21</sup>

A 2018 Hayes review of 16 qualifying studies found that CH testing has some capacity to detect presence of glaucoma, to predict risk for glaucoma progression, and to predict response of glaucoma to certain types of treatment.<sup>22</sup> However, “the evidence is mainly comprised of very poor quality correlation studies, which lack the rigor to determine diagnostic or prognostic accuracy. Most of these studies did not use reliable methods to determine the accuracy of diagnosis or prognosis. No studies were identified that directly assessed the clinical utility of CH measurement for selecting treatment for glaucoma or for impacting long-term health outcomes.” This perspective was also reflected in the one recent meta-analysis; although CH was correlated with glaucoma, there were no findings related to the use of CH for the prognosis, diagnosis, or management of glaucoma.<sup>12</sup>

In summary, CH is promising as a risk assessment tool in the diagnosis and management of glaucoma or corneal pathology. However, while the body of evidence is large, the overall quality is low. The studies are relatively small, observational, often confounded by lack of treatment control, uniformly citing simple correlations, precluding cause-and-effect conclusions. Not only are there no Level I studies, none of the reviewed studies demonstrate that CH measurement alters clinical management and improves clinical outcomes. A wide array of tests are accepted for detection and monitoring of glaucoma (tonometry for IOP, perimetry to assess visual field, ophthalmoscopy to detect a glaucomatous optic nerve head (ONH) and RNFL changes, and pachymetry for CCT). It is still unclear whether CH provides useful additional information, much less its optimal role in any diagnostic, prognostic, and treatment algorithm. Randomized controlled trials (RCTs) that compare outcomes in patients whose treatment is selected based on CH are needed to determine definitive patient selection criteria and clinical utility. The lack of level I evidence, absence of proven clinical utility, no clinical practice guideline endorsement,<sup>18-20,23</sup> as well as neither Medicare<sup>24-25</sup> nor commercial<sup>26-28</sup> coverage, argue strongly against current CH coverage as reasonable and necessary for treatment of Medicare patients.

---

## General Information

### Associated Information

N/A

### Sources of Information

N/A

### Bibliography

1. FDA 510K Ocular Response Analyzer clearance. 2004; [https://www.accessdata.fda.gov/cdrh\\_docs/pdf3/K032799.pdf](https://www.accessdata.fda.gov/cdrh_docs/pdf3/K032799.pdf).
2. Congdon NG, Broman AT, Bandeen-Roche K, Grover D, Quigley HA. Central corneal thickness and corneal hysteresis associated with glaucoma damage. *Am J Ophthalmol*. 2006;141(5):868-875.
3. Narayanaswamy A, Su DH, Baskaran M, et al. Comparison of ocular response analyzer parameters in chinese subjects with primary angle-closure and primary open-angle glaucoma. *Arch Ophthalmol*. 2011;129(4):429-434.
4. Agarwal DR, Ehrlich JR, Shimmyo M, Radcliffe NM. The relationship between corneal hysteresis and the magnitude of intraocular pressure reduction with topical prostaglandin therapy. *Br J Ophthalmol*.

- 2012;96(2):254-257.
5. De Moraes CV, Hill V, Tello C, Liebmann JM, Ritch R. Lower corneal hysteresis is associated with more rapid glaucomatous visual field progression. *J Glaucoma*. 2012;21(4):209-213.
  6. Vu DM, Silva FQ, Haseltine SJ, Ehrlich JR, Radcliffe NM. Relationship between corneal hysteresis and optic nerve parameters measured with spectral domain optical coherence tomography. *Graefes Arch Clin Exp Ophthalmol*. 2013;251(7):1777-1783.
  7. Murphy ML, Pokrovskaya O, Galligan M, O'Brien C. Corneal hysteresis in patients with glaucoma-like optic discs, ocular hypertension and glaucoma. *BMC Ophthalmol*. 2017;17(1):1.
  8. Mansouri K, Leite MT, Weinreb RN, Tafreshi A, Zangwill LM, Medeiros FA. Association between corneal biomechanical properties and glaucoma severity. *Am J Ophthalmol*. 2012;153(3):419-427 e411.
  9. Medeiros FA, Meira-Freitas D, Lisboa R, Kuang TM, Zangwill LM, Weinreb RN. Corneal hysteresis as a risk factor for glaucoma progression: a prospective longitudinal study. *Ophthalmology*. 2013;120(8):1533-1540.
  10. Zhang C, Tatham AJ, Abe RY, et al. Corneal Hysteresis and Progressive Retinal Nerve Fiber Layer Loss in Glaucoma. *Am J Ophthalmol*. 2016;166:29-36.
  11. Susanna CN, Susanna BN, Diniz-Filho A, et al. A Prospective Longitudinal Study to Investigate Corneal Hysteresis as a Risk Factor for Predicting Development of Glaucoma. *Am J Ophthalmol*. 2018.
  12. Gaspar R, Pinto LA, Sousa DC. Corneal properties and glaucoma: a review of the literature and meta-analysis. *Arq Bras Oftalmol*. 2017;80(3):202-206.
  13. Chen M, Kueny L, Schwartz AL. The role of corneal hysteresis during the evaluation of patients with possible normal-tension glaucoma. *Clin Ophthalmol*. 2018;12:555-559.
  14. Shah S, Laiquzzaman M, Bhojwani R, Mantry S, Cunliffe I. Assessment of the biomechanical properties of the cornea with the ocular response analyzer in normal and keratoconic eyes. *Invest Ophthalmol Vis Sci*. 2007;48(7):3026-3031.
  15. Fontes BM, Ambrosio R, Jr., Jardim D, Velarde GC, Nose W. Corneal biomechanical metrics and anterior segment parameters in mild keratoconus. *Ophthalmology*. 2010;117(4):673-679.
  16. Touboul D, Benard A, Mahmoud AM, Gallois A, Colin J, Roberts CJ. Early biomechanical keratoconus pattern measured with an ocular response analyzer: curve analysis. *J Cataract Refract Surg*. 2011;37(12):2144-2150.
  17. AAO Glaucoma Summary Benchmarks. 2017; <https://www.aao.org/summary-benchmark-detail/glaucoma-summary-benchmarks-2020>
  18. Prum BE, Jr., Lim MC, Mansberger SL, et al. Primary Open-Angle Glaucoma Suspect Preferred Practice Pattern((R)) Guidelines. *Ophthalmology*. 2016;123(1):P112-151.
  19. Prum BE, Jr., Rosenberg LF, Gedde SJ, et al. Primary Open-Angle Glaucoma Preferred Practice Pattern((R)) Guidelines. *Ophthalmology*. 2016;123(1):P41-P111.
  20. Canadian Association of Optometrists (CAO). Screening, Diagnosis, and Management of Open Angle Glaucoma: An Evidence-Based Guideline for Canadian Optometrists. 2017; <https://openjournals.uwaterloo.ca/index.php/cjo/article/view/1606>
  21. Garcia-Ferrer FJ, Akpek EK, Amescua G, et al. Corneal Ectasia Preferred Practice Pattern(R). *Ophthalmology*. 2018.
  22. Hayes: Measurement of Corneal Hysteresis for the Diagnosis and Management of Glaucoma 2018; <https://www.hayesinc.com/publications/evidence-analysis/health-technology-assessment/measurement-of-corneal-hysteresis-for-the-diagnosis-and-management-of-glaucoma/dir-cornealhysteresis4402/>
  23. NICE: ORA G3 to measure corneal hysteresis. 2018; <https://www.nice.org.uk/advice/mib150>
  24. Palmetto GBA: Corneal Hysteresis (L38211)
  25. WPS: Corneal Hysteresis (A56910).
  26. UnitedHealthcare: CORNEAL HYSTERESIS AND INTRAOCULAR PRESSURE MEASUREMENT (Policy Number: 2023T0133Z). [UHC Corneal Hysteresis](https://www.uhc.com/healthcare/corneal-hysteresis).
  27. Aetna: Glaucoma Testing Number: 0622. [http://www.aetna.com/cpb/medical/data/600\\_699/0622.html](http://www.aetna.com/cpb/medical/data/600_699/0622.html).
  28. BCBSMA: Medical Technology Assessment Investigational (Non-Covered) Services List.



# Revision History Information

REVISION HISTORY DATE	REVISION HISTORY NUMBER	REVISION HISTORY EXPLANATION	REASONS FOR CHANGE
07/27/2023	R4	Posted 07/27/2023: Biannual review completed. Minor grammatical changes made throughout. Updated references to include superscripts throughout the LCD. The following references were deleted as they have been retired: First Coast Service Options: Noncovered Services (L33777), Noridian Non-Covered Services (L35008) and Palmetto GBA: Noncovered Services other than CPT® Category III Noncovered Services (L36954). Updated reference 24 to LCD for Palmetto GBA: Corneal Hysteresis (L38211). Updated reference, WPS Corneal Hysteresis (A54957) to current LCA, WPS: Corneal Hysteresis (A56910). Citations were corrected throughout the LCD to reflect the revised bibliography. Broken hyperlinks in the LCD were updated. No change in coverage.	<ul style="list-style-type: none"> <li>• Typographical Error</li> </ul>
09/30/2021	R3	02/01/2022 - Removed broken link in #30 of the bibliography. Added correct links to #17, #20, and #22 of the bibliography.	<ul style="list-style-type: none"> <li>• Typographical Error</li> </ul>
09/30/2021	R2	09/30/2021 Review completed 08/30/2021 with no change in coverage. Minor grammatical corrections made throughout the LCD.	<ul style="list-style-type: none"> <li>• Other (Review)</li> </ul>
11/01/2019	R1	Change Request 10901 Local Coverage Determinations (LCDs): it will no longer be appropriate to include Current Procedure Terminology (CPT)/Health Care Procedure Coding System (HCPCS) codes or International Classification of Diseases Tenth Revision-Clinical Modification (ICD-10-CM) codes in the LCDs. All CPT/HCPCS, ICD-10 codes, and Billing and Coding Guidelines have been removed from this LCD and placed in the Billing and Coding Article related to this LCD. Consistent with Change Request 10901, if any language from IOMs and/or regulations was present in the LCD, it has been removed and the applicable manual/regulation has been referenced.	<ul style="list-style-type: none"> <li>• Revisions Due To Code Removal</li> </ul>

## Associated Documents

### Attachments

N/A

### Related Local Coverage Documents

### Articles

[A56910 - Billing and Coding: Corneal Hysteresis](#)

[A56917 - Response to Comments: Corneal Hysteresis L38211](#)

### LCDs

[DL38211 - \(MCD Archive Site\)](#)

### Related National Coverage Documents

N/A

### Public Versions

UPDATED ON	EFFECTIVE DATES	STATUS
07/19/2023	07/27/2023 - N/A	Currently in Effect (This Version)
02/02/2022	09/30/2021 - 07/26/2023	Superseded
09/20/2021	09/30/2021 - N/A	Superseded

Some older versions have been archived. Please visit the MCD Archive Site to retrieve them.

---

## Keywords

- Corneal
- Hysteresis